

## **BabiTENS Product Specification**

<b>Name:</b>	BabiTENS – Obstetric Transcutaneous Electrical Nerve Stimulator
<b>Type of Equipment:</b>	Internally powered Type BF conforms to B.S. 5724
<b>Channels:</b>	Dual
<b>Amplitude:</b>	0-70 milli Amps (35V), 120mA peak to peak (60V), +/- 15%, 500 Ohm Load
<b>Power Source:</b>	9 Volt Battery Eveready 1604A or 1604D Ni/Cad: Varta V 7/8R or TR 7/8
<b>Pulse Width:</b>	40-250 micro seconds +/- 15%
<b>Rates:</b>	2-200 Hz, +/- 15%, +/- 1 Hz. At the lowest rate
<b>Scalling:</b>	Low/High for rate and width
<b>Burst Rates:</b>	Two bursts per second
<b>Burst Frequency:</b>	Adjustable; controlled by rate dial
<b>Mode Function:</b>	Normally burst; continuous when switch is depressed
<b>Function Switch:</b>	Remote hand switch attached to top of unit. Click On. Click Off